



Statement of Gamma Irradiation Treatment on Equflow PVDF Turbine Flowmeters

Manufacturer:

Equflow bv
PO BOX 6671
NL-6503 GD Nijmegen
Netherlands

Declares that:

Description	PVDF Turbine Tubes 0045 and 0085
Article P/N	0045.P.H.0.00.CX 0045.P.H.0.00.TX 0085.P.H.0.00.TX 0085.P.H.0.00.CX
Prod. number	N.A.
Date of manufact.	N.A.
Expire	N.A.
Serial number	N.A.
	Sterilized with 50 kGy Gamma Radiation

Did not show any significant change in the accuracy and repeatability of the Flowmeters, before and after Gamma Sterilization

Date: 1 March 2010

.....
Released by (signature)
Name: Fred Wooninck

Manager Production